

Room Name	Neonatal Bay - Low Dependency Care		Room Number NBLD
Revision	4		
Issue Date	13.05.2020		
BRIEFING			
Briefed Area	12.50 m ²		
Hours of Operation	24 Hours		
Occupancy	1 baby; 1-2 staff; 1-2 carers		
Acoustics	Sound attenuation level high		
Remarks			
Description / Special Requirements	A single bay for neonates requiring low dependency nursing and medical treatment, that may include oxygen therapy, intravenous therapy and continuous cardiorespiratory monitoring. The bay will have ready access to a Type A Handwashing Basin, shared between bays. Facilities for storing parent/carer's belongings is required and access to baby bathing. An external outlook is desirable from a group of neonatal bays. Lighting will require special consideration and should be dimmable, zoned in parent spaces and direct overhead lighting should be avoided.		
SERVICES			
General Requirements	BODY PROTECTED	<input checked="" type="checkbox"/>	CARDIAC PROTECTED <input type="checkbox"/>
Air	AIRCONDITIONING	<input checked="" type="checkbox"/>	AIRCONDITIONING: hepa filtered <input type="checkbox"/>
	AIRCONDITIONING: positive pressure	<input type="checkbox"/>	AIRCONDITIONING: negative pressure <input type="checkbox"/>
	EXHAUST: room exhaust	<input type="checkbox"/>	VENTILATION <input type="checkbox"/>
Lighting	LIGHTING: colour corrected	<input checked="" type="checkbox"/>	LIGHTING: dimmable <input checked="" type="checkbox"/>
	LIGHTING: downlights	<input checked="" type="checkbox"/>	LIGHTING: downlights, dimmable <input type="checkbox"/>
	LIGHTING: fluorescent/LED, downlights	<input type="checkbox"/>	LIGHTING: fluorescent/LED, flush with ceiling, tamper proof <input type="checkbox"/>
	LIGHTING: fluorescent/LED, general	<input type="checkbox"/>	LIGHTING: indirect <input type="checkbox"/>
Nurse Call	NURSE CALL SYSTEM	<input checked="" type="checkbox"/>	

Room Fabrics

AusHFG code	Description	Comment
CLFS-006	CEILING: plasterboard, flush set, suspended, paint, washable	consider acoustic properties
CLCN-008	CORNICE: square set flush between wall and ceiling, paint	
FLVY-007	FLOOR FINISH: vinyl, seamless, coved, standard slip resistant	low VOC vinyl, consider acoustic options
FLSK-010	SKIRTING: vinyl, floor vinyl coved, 150H, prefinished	
WLWA-004	WALL FINISH: paint, acrylic, washable	
WLWA-025	WALL FINISH: vinyl, to 1200 AFFL	
WLPR-006	WALL PROTECTION: corner guards	

Fittings, Furniture and Equipment (FF&E) Items

AusHFG code	Description	Group	Qty	Comment
MMGE-002	ADAPTER: suction	3	3	on services panel
FQCW-043	BIN: waste, sharps, on trolley	3	1	mounted on trolley FQSM-035
FIHR-067	BRACKET: for monitor, wall mounted, articulated, adj height	2	1	
FIHR-083	BRACKET: for neonatal camera, crib mounted	2	1	optional
FIHR-084	BRACKET: for oxygen and air mixer	2	1	

Room Name | Neonatal Bay - Low Dependency Care
Room Number NBLD

AusHFG code	Description	Group	Qty	Comment
FIHR-016	BRACKET: for sharps bin	2	1	mounted on trolley FQSM-035
FIHR-019	BRACKET: for suction bottle	2	3	
ITCL-036	CAMERA: neonatal	3	1	optional; must be able to be turned off in the bay; storage and use of images to be considered
FQBS-019	CHAIR: office, ergonomic, high	3	1	optional
FQBS-055	CHAIR: overnight fold out	3	1	compatible with kangaroo care, breastfeeding
FQBS-028	CHAIR: visitor, clinical areas	3	1	
ITNE-007	COMPUTER: desktop, with display screen	3	1	optional; on trolley
FQSM-044	CUPBOARD: mobile, double door, adjustable shelves	3	1	
FIDI-001	DISPENSER: alcohol based hand rub	2	2	shared
MGFP-002	FLOWMETER: medical air	3	2	on services panel
MGFP-004	FLOWMETER: oxygen	3	2	on services panel
FQGE-091	FRIDGE: milk, small	3	1	optional
MMMA-007	INCUBATOR: infant, humidicrib	3	1	or bassinette
LIFX-013	LIGHT: procedure, ceiling mounted, small	1	1	
MMMA-018	MIXER: oxygen and air	1	1	
MMHA-005	MONITOR: patient, cardiac	3	1	
MMGE-076	PUMP: breast	3	1	on trolley/stand
FIJO-194	SHELF: laminate	1	1	optional; for small milk fridge FQGE-091
MMGE-100	SUCTION BOTTLE	3	3	
ITNE-050	TELEPHONE: handset, desktop, handsfree	3	1	
FQSM-035	TROLLEY: small	3	2	
FQCA-086	WARMER: bottle, mobile	3	1	
FIBM-018	WHITEBOARD: fixed	2	1	may be optionally replaced with a care board

Engineering Services

AusHFG code	Description	Group	Qty	Comment
ITCL-004	BUTTON: nurse call, emergency	1	1	on services panel
ITCL-006	BUTTON: nurse call, patient to staff	1	1	on wall, for parent/carer use
ITCL-008	BUTTON: nurse call, staff assist	1	1	on services panel
ELGP-208	GPO: double, wall mounted	1	1	to parent/carer area; may include optional USB charging point
ELGP-223	GPO: emergency power, double, wall mounted	1	4	
ELGP-122.1	GPO: emergency power, single, on services panel (V)	1	6	on services panel
ELGP-123	GPO: emergency power, single, wall mounted	1	2	
ELGP-101	GPO: single, ceiling mounted	1	1	
ELGP-110	GPO: single, wall mounted, cleaner	1	1	

Room Name | Neonatal Bay - Low Dependency Care

Room Number NBLD

AusHFG code	Description	Group	Qty	Comment
ELGP-133	GPO: UPS, single, on services panel	1	2	on services panel
ITCL-028	HANDSET: nurse call, with patient to staff call button	1	1	on wall, for parent/carer use
LIFX-017	LIGHT: night light	1	1	switch to be located at cubicle or staff station
ELPR-004	MONITOR: line isolation	1	1	
ITIN-025	OUTLET: data, double RJ45, on services panel	1	2	on services panel
ITIN-026	OUTLET: data, double RJ45, wall mounted	1	1	
ITIN-016	OUTLET: data, single RJ45, wall mounted	1	1	
MGAS-022	OUTLET: medical air (MA), on services panel	1	3	on services panel, 1 no. optional
MGAS-042	OUTLET: oxygen (O2), on services panel	1	3	on services panel, 1 no. optional
MGAS-062	OUTLET: suction, on services panel	1	3	on services panel
ELPR-006	RCD: residual current device	1	2	on services panel; Line Isolation Monitor may be provided for UPS
MMSP-026	SERVICES PANEL: wall mounted	1	3	flush mounted
ELBO-012	SWITCH: light, dimmer	1	1	
ELPR-007	UPS: status indicator	1	1	on services panel