

Room Name	Operating Room - General, 42m2		Room Number ORGN
Revision	4		
Issue Date	20.03.17		
BRIEFING			
Briefed Area	42.00 m ²		
Hours of Operation	24 Hours		
Occupancy	1 patient; 5 - 8 staff		
Acoustics			
Remarks			
Description / Special Requirements	The Operating Room - General provides an aseptic environment in which to carry out surgical procedures under local, regional or general anaesthetic. The General Operating Room may be used for General surgery and specialties including ENT, Urology, Gynaecology, Ophthalmology, Plastic Surgery and any other procedures that do not require bulky equipment. The Operating Room will have direct access to a Scrub-up Room, exit bay and sterile stock/set-up area. Additional Design Considerations: Minimum critical dimension 6000mm. 'Laser In Use' warning lights may be required outside all exit doors to room if laser equipment is to be used. Lead shielding may be required depending on use of X-ray equipment - to be assessed by Radiation Consultant. External windows are desirable - double glazed with integral venetian blinds.		
SERVICES			
General Requirements	BODY PROTECTED = No	CARDIAC PROTECTED = Yes	
Air	AIRCONDITIONING = No AIRCONDITIONING: positive pressure = Yes EXHAUST: room exhaust = No	AIRCONDITIONING: hepa filtered = Yes AIRCONDITIONING: negative pressure = No VENTILATION = No	
Lighting	LIGHTING: colour corrected = Yes LIGHTING: downlights = No LIGHTING: fluorescent, downlights = No LIGHTING: fluorescent, general = No	LIGHTING: dimmable = No LIGHTING: downlights, dimmable = No LIGHTING: fluorescent, flush with ceiling, tamper proof = No LIGHTING: indirect = No	
Nurse Call	NURSE CALL SYSTEM = Yes		

Room Fabrics

AusHFG code	Description	Comment
CLFS-006	CEILING: plasterboard, flush set, suspended, paint, washable	
CLCN-007	CORNICE: plasterboard, flush coved between wall and ceiling, paint	
DWGL-020	DOOR: 1100mm c/o, single leaf, observation panel, solid core/glass, paint	1100mm clear opening, to clean-up room, quantity to suit layout
DWGL-042	DOOR: 1500mm c/o, 1 1/2 leaf, observation panel, solid core/glass, paint	1500mm clear opening to patient entry/exit, quantity to suit layout
DWGL-005	DOOR: 910mm c/o, single leaf, observation panel, solid core/glass, paint	910mm clear opening to set-up/scrub rooms, quantity to suit layout
DWPR-006	DOOR PROTECTION: protection plate, to 900 AFFL, PVC, prefinished	Equal to Acrovyn
DWPR-010	DOOR PROTECTION: to door frame, PVC, prefinished	Equal to Acrovyn
FLVY-007	FLOOR FINISH: vinyl, seamless, coved, standard slip resistant	
DWOB-003	OBSERVATION: internal window, alum/glass, double glazed, integral venetian blind, powdercoat	optional
DWOB-010	OBSERVATION: through door/s, glass, clear	with cover (sliding)

Room Name | **Operating Room - General, 42m2**
Room Number ORGN

AusHFG code	Description	Comment
FLSK-010	SKIRTING: vinyl, floor vinyl coved, 150H, prefinished	
WLWA-005	WALL FINISH: paint, scrub resistant	
WLWA-022	WALL FINISH: vinyl, to 900 AFFL	

Fittings, Furniture and Equipment (FF&E) Items

AusHFG code	Description	Group	Qty	Comment
MMGE-002	ADAPTER: suction	3	4	to services pendants
FIJO-022	BENCH: 600D, open under, resinate finish, mobile	3	1	for computer, optional
DWWD-007	BLIND: venetian, integral	1	1	to internal window, if provided
FIHR-006	BRACKET: for display screen, wall mounted	1	1	optional, for TV display screen if required
FIHR-019	BRACKET: for suction bottle	2	4	
MMHA-001	CARRIER: suction bottle	3	1	mobile
MMGE-014	CHAIR: anaesthetist	3	1	
ITCL-016	CLOCK: clock system slave, analogue	1	1	recessed
ITCL-022	CLOCK: clock system slave, elapsed time	1	1	beside analogue, recessed
ITNE-007	COMPUTER: desktop, with display screen	3	1	optional
FIDI-014	DISPENSER: disposable glove	2	1	
ITBU-011	DISPLAY SCREEN: MATV	3	1	for television, could be replaced with IPTV, to be decided at project level
MGFP-002	FLOWMETER: medical air	3	2	to services pendants
MGFP-004	FLOWMETER: oxygen	3	4	to services pendants
MMGE-045	IV STAND: mobile	3	2	
MMOP-028	KICKBUCKET & STAND:	3	2	
MMSP-028	LIGHT: operating, 2 arm, 950/550 diam nom.	1	1	
MMGE-048	LINEN CARRIER: dirty, single	3	1	
MMOP-040	MACHINE: anaesthetic	3	1	
MMOP-043	MACHINE: electrosurgical, diathermy	3	1	
MMHA-004	MONITOR: patient, anaesthetic	3	1	with printer, on anaesth machine
MMSP-015	PENDANT: single, anaesthetic (fixed or adjustable height)	1	1	cable support arm optional
MMSP-016	PENDANT: single, equipment (fixed or adjustable height)	1	1	surgical, cable support arm optional
MMOP-069	STAND: bowl, single	3	2	
MMOP-071	STOOL: operating, mobile	3	2	
MMGE-100	SUCTION BOTTLE	3	4	in mobile carrier
MMOP-076	TABLE: mayo	3	1	
MMOP-079	TABLE: operating, manual	3	1	
ITNE-050	TELEPHONE: handset, desktop, handsfree	3	1	to bench
MMOP-084	TOURNIQUET: surgical	3	1	

Room Name | **Operating Room - General, 42m2**
Room Number ORGN

AusHFG code	Description	Group	Qty	Comment
MMOP-085	TROLLEY: anaesthetic	3	1	
MMGE-122	TROLLEY: gowns	3	1	
MMGE-123	TROLLEY: instrument	3	4	
MMGE-156	VIEWER: X-ray, 4 screen	1	1	recessed, 4x1 layout
FIBM-019	WHITEBOARD: fixed, magnetic	2	1	900x900, optional

Engineering Services

AusHFG code	Description	Group	Qty	Comment
MGFP-001	ALARM PANEL: medical gases	1	1	to services panel P1
ITCL-004	BUTTON: nurse call, emergency	1	1	to services panel P4, + indicator button & light
ITCL-005	BUTTON: nurse call, orderly / porter	1	1	to services panel P4, optional
ITCL-008	BUTTON: nurse call, staff assist	1	1	to services panel P4, + indicator button & light
ELBO-003	CONTROL PANEL: lighting	1	1	to services panel P2
ITCL-023	CONTROL PANEL: master clock system	1	1	to services panel P2
MGFP-007	CONTROL PANEL: tool air	1	1	to services panel P1
ELGE-001	DIRECT CONNECTION: power	1	3	XRV, clocks, lights
ELBO-005	EQUIPOTENTIAL JUNCTION BOX	1	1	to services panel P1
ELGP-209	GPO: double, wall mounted, RCD protected	1	1	to bench, optional
ELGP-121	GPO: emergency power, single, on pendant	1	8	4 per pendant
ELGP-122	GPO: emergency power, single, on services panel	1	3	to services panel P1
ELGP-123	GPO: emergency power, single, wall mounted	1	1	to wall TV display screen. optional
ELGP-104	GPO: single, on pendant	1	8	4 per pendant
ELGP-105	GPO: single, on services panel	1	2	to services panel P3, 1 no. 20 amp for laser (optional)
ELGP-105	GPO: single, on services panel	1	2	to services panel P2
ELGP-110	GPO: single, wall mounted, cleaner	1	1	
LIFX-003	LIGHT: illuminated sign, "Laser in Use"	1	4	outside room entries, if laser to be used
ITIN-024	OUTLET: data, double RJ45, on pendant	1	2	1 to each pendant
ITIN-026	OUTLET: data, double RJ45, wall mounted	1	1	to bench; 1 outlet for telephone
ITIN-016	OUTLET: data, single RJ45, wall mounted	1	1	to wall TV display screen, optional, with monitor connection facilities and cabling
MGAS-021	OUTLET: medical air (MA), on pendant	1	2	1 on each pendant
MGAS-031	OUTLET: nitrous oxide (N2O), on pendant	1	2	1 on each pendant
MGAS-041	OUTLET: oxygen (O2), on pendant	1	4	2 on each pendant
MGAS-052	OUTLET: scavenge, on pendant, with adapter	1	2	1 on each pendant
MGAS-061	OUTLET: suction, on pendant	1	4	2 on each pendant

Room Name | **Operating Room - General, 42m2**
Room Number ORGN

AusHFG code	Description	Group	Qty	Comment
MGAS-074	OUTLET: tool air, on pendant	1	2	1 on each pendant, controls on wall mounted services panel
MGAS-080	OUTLET: tourniquet air, with controls	1	1	to services panel P1, and controls, optional
ELPR-006	RCD: residual current device	1	3	to services panel P1, P2 and P3
MMSP-026	SERVICES PANEL: wall mounted	2	4	flush mounted
MESU-013	SUPPLY AIR: grille, HEPA filtered	1	4	low level
ELBO-009	SWITCH: "Laser in use"	1	1	to services panel P4