



# AUSTRALASIAN HFG STANDARD COMPONENTS

Room Layout Sheet


RLS Room Code: **LINAC-CR**




Room Name: **LINEAR ACCELERATOR CONTROL ROOM**




CODE	DESCRIPTION
DWPR-001	DOOR PROTECTION: kickplate, to 300 AFFL
ELBO-001	BUTTON: emergency stop
ELBO-012	SWITCH: light, dimmer
ELGP-106	GPO: single, special, for equipment
ELGP-208	GPO: double, wall mounted
ELGP-223	GPO: emergency power, double, wall mounted
ELPR-001	BODY PROTECTED ELECTRICAL AREA
ELPR-006	RCD: residual current device
FIBM-019	WHITEBOARD: fixed, magnetic
FIHR-067	BRACKET: for monitor, wall mounted, articulated, adj height
FIHR-080	BRACKET: for bench, 750D, wall mounted
FUJO-031	BENCH: 750D, open under, laminate finish
FUJO-099	CAC: cable access cap
FUJO-145	CUPBOARD: wall mount, double door, adj shelves, lockable
FUJO-212	BULKHEAD: joinery
FUJO-213	PANEL: joinery carcass end
FLSK-010	SKIRTING: vinyl, floor vinyl coved, 150H, prefinished
FLVY-007	FLOOR FINISH: vinyl, seamless, coved, standard slip resistant
FOBS-018	CHAIR: office, ergonomic
FOGE-005	CLOCK: analogue, battery operated
ITAV-014	MICROPHONE: surface mounted, gooseneck
ITAV-057	SPEAKER: ceiling recessed
ITCL-004	BUTTON: nurse call, emergency
ITCL-008	BUTTON: nurse call, staff assist
ITIN-026	OUTLET: data, double RJ45, wall mounted
ITIN-036	OUTLET: data, triple RJ45, wall mounted
ITIN-055	OUTLET: data, USB, wall mounted
ITNE-049	TELEPHONE: handset, desktop
ITNE-067	KEYBOARD: keyboard and mouse only
ITNE-071	COMPUTER: desktop, with dual display screens
ITNE-074	DISPLAY SCREEN: patient observation
ITSE-006	BUTTON: duress, fixed
MMRA-006	ACCESS PANEL: medical physics port
MMRA-007	CABINET: linear accelerator computer cabinet
MMRA-009	COMPUTER: linear accelerator control system PC
MMRA-010	CONSOLE: control, linear accelerator
WLWA-004	WALL FINISH: paint, acrylic, washable




PLAN




ELEVATION 1



ELEVATION 3



ELEVATION 4



ELEVATION 2

REV	DESCRIPTION	DATE
1	First Issue	28.06.2021

FOR GUIDANCE

SCALE	RLS Room Code	REV
1 : 50 @ A3	LINAC-CR	1

